

CERTIFIED ORIGINAL



HOLLOWAY HOUSTON, INC.

LEADERS IN LIFTING SAFETY
World's Largest CROSBY Distributor
5833 ARMOUR DR. HOUSTON, TX 77020
PH#713-674-5631 FX#713-674-8352

Certificate of Calibration

Client: Bastrop Scale Co Inc.
Device Calibrated: 150te Compression Load Cell
Manufacturer: Coti
Model #: CG26S5
Type: Cabled Remote
Device calibrated S/N: 2011027002
Indicator(s): GSE 250 Digtal Display

Invoice#: 1070133

Serial #: 2011027002

Calibration Units: Lbf
Date Calibrated: November 1, 2018
Verification Due: November 1, 2019
Method of calibration: Compression

Temperature: 73F*
***Correction:** n/a

Calibration Factor: n/a

This device meets the accuracy standards of ASTM E4-16 (+/- 1% of reading) as indicated below.
Any other loading data is shown for reference only.

Reference True Load Lbf	As Found Reading Lbf	Run 1 Indication Lbf	Percent Error of Reading	Meets ASTM E-4 Accuracy?	Run 2 Indication Lbf	Percent Error of Reading	Meets ASTM E-4 Accuracy?	Repeatability as a percent
0		0			0			
30,000	30,750	30,200	0.67	YES	30,200	0.67	YES	0.00
50,000	51,050	50,200	0.40	YES	50,200	0.40	YES	0.00
100,000	101,950	100,200	0.20	YES	100,200	0.20	YES	0.00
150,000	152,850	150,400	0.27	YES	150,400	0.27	YES	0.00
200,000	203,750	200,500	0.25	YES	200,500	0.25	YES	0.00
300,000	305,050	300,000	0.00	YES	300,000	0.00	YES	0.00

Adjustment made to meet specified accuracy?

HHI reference standard(s):

Serial	Calibrated	Date Due	Class A loading range	Verified by	MFG by	Temperature
B263412	11/7/2017	11/7/2019	17,500- 350,000 Lbf	NIST	Coti	23C

Calibration By: 

Date: 11-1-18

Approved By: 

Date: 11-1-18